



## RUDADERM®-universal

REF 090094

### Product Description, Intended use,

#### Application

The plaster stripes consist of a carrier material (0.9 x 3.8cm) which is made of a flexible, beige, waterproof synthetic and is coated with a hypoallergenic adhesive. The strips are individually packed and are used to cover injection sites, small punctures (for example vein punctures) and primary healing smaller injuries.

#### Composition

Polyethylene, vinyl acetate - carrier  
Viscose, polyester, polypropylene,  
polyethylene – wound pad,  
hypoallergenic adhesive

#### Incident reporting

According to MDR (EU) 2017/745, if serious incidents occur in relation to the device, they must be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

#### Normative and Regulative

#### Requirements, Common Standards

Medical Device according to MDD 93/42/EEC,  
MDR (EU) 2017/745.

The product does not contain dangerous toxic substances according to REACH.

### Packaging

Primary packaging: Paper packaging

Secondary packaging: folding box  
made of cellulose

Tertiary packaging: Carton made of  
Cellulose

### Symbols used in labelling

Explanations at [www.nobamed.com](http://www.nobamed.com)



STERILE EO



CE  
0123

Marking on all packaging levels with CE and according to DIN EN ISO 15223-1- and DIN EN 1041.

### Storage and Transport

Dry and dustfree.

### Sterile device

Before using a sterile product, visually inspect the packaging to ensure that it is intact.

### Single use device

Reusing a single use medical device can lead to microbiological danger. Reprocessing for reuse can decrease the product's performance significantly.

### Disposal

According to locally applicable legal regulations and standards of infection prophylaxis.