



## NOBADRAPE® Variable Fenestrated Drape

REF 764075

### Product Description, Intended use, Application

The 2-ply blue drape (75 cm x 90 cm) consists of two equal parts which can be arranged variably so that the resulting fenestration complies exactly with the user's requirements. Both parts are sterile packed and belong together as one

product. The drape is used for the sterile demarcation of the operation field. The top layer consists of an absorbent polypropylene nonwoven. The bottom layer consists of a moisture-impermeable polyethylene film providing a reliable barrier protection.

Easily removable control stickers on the packaging facilitate surgical documentation.

### Composition

Polypropylene, polyethylene, polyester, hypopallergenic adhesive

### Contraindications

The product should not be used in the case of a known allergy against the material.

### Normative and Regulative

#### Requirements, Common Standards

Medical Device according to *MDD 93/42/EEC, MDR (EU) 2017/745*.

The product complies with the requirements of the **DIN EN 13795-1**, "Surgical clothing and drapes – Requirements and test methods – Part 1: Surgical drapes and gowns (requirements of the performance level 'high' and 'standard').

Sterilization of the product complies with DIN EN 11135.

The product does not contain dangerous toxic substances according to REACH.

### Packaging

|            |                               |
|------------|-------------------------------|
| Primary:   | paper-film packaging          |
| Secondary: | folding box made of cellulose |
| Tertiary:  | carton made of cellulose      |

### Symbols used in labelling

Explanation at [www.nobamed.com](http://www.nobamed.com)



Marking on all packaging levels with CE and according to DIN EN ISO 15223-1- and DIN EN ISO 20417

### Storage and transport

To be stored in a dry and dust-free environment, protect from sunlight.

### Sterile device

Before using a sterile product, visually inspect the packaging to ensure that it is intact.

### Single use device

Reusing a single use medical device can lead to microbiological danger. Reprocessing for reuse can decrease the product's performance significantly.

### Disposal

According to locally applicable legal regulations and standards of infection prophylaxis.